

Myocardial strain imaging: review of general principles, validation, and sources of discrepancies

M.S. Amzulescu¹, M. De Craene², H. Langet³, A. Pasquet¹, D. Vancraeynest¹, A.C. Pouleur¹, J.L. Vanoverschelde¹, and B.L. Gerber^{1*}

¹Division of Cardiology, Department of Cardiovascular Diseases, Cliniques Universitaires St. Luc, Pôle de Recherche Cardiovasculaire (CARD), Institut de Recherche Expérimentale et Clinique (IREC), Université Catholique de Louvain, Av Hippocrate 10/2806, B-1200 Brussels, Belgium; ²Philips Research, Medical Imaging (Medisys), 33 rue de Verdun, CS60055, 92156 Suresnes Cedex, France; and ³Clinical Research Board, Philips Research, 33 rue de Verdun, CS60055, 92156 Suresnes Cedex, France

Received 6 February 2019; editorial decision 21 February 2019; accepted 7 March 2019

Myocardial tissue tracking imaging techniques have been developed for a more accurate evaluation of myocardial deformation (i.e. strain), with the potential to overcome the limitations of ejection fraction (EF) and to contribute, incremental to EF, to the diagnosis and prognosis in cardiac diseases. While most of the deformation imaging techniques are based on the similar principles of detecting and tracking specific patterns within an image, there are intra- and inter-imaging modality inconsistencies limiting the wide clinical applicability of strain. In this review, we aimed to describe the particularities of the echocardiographic and cardiac magnetic resonance deformation techniques, in order to understand the discrepancies in strain measurement, focusing on the potential sources of variation: related to the software used to analyse the data, to the different physics of image acquisition and the different principles of 2D vs. 3D approaches. As strain measurements are not interchangeable, it is highly desirable to work with validated strain assessment tools, in order to derive information from evidence-based data. There is, however, a lack of solid validation of the current tissue tracking techniques, as only a few of the commercial deformation imaging softwares have been properly investigated. We have, therefore, addressed in this review the neglected issue of sub-optimal validation of tissue tracking techniques, in order to advocate for this matter.

Keywords

strain • speckle tracking imaging • feature tracking • tagging • echocardiography • cMR • review

Introduction

Assessment of cardiac contractile function remains a challenge in current cardiology. Indeed, ejection fraction (EF), the traditional parameter used to describe left ventricular (LV) function, presents significant limitations,¹ related to its volumetric nature, suboptimal reproducibility, and inability to reflect regional LV function. This has prompted for a more in-depth characterization of LV mechanics through non-invasive evaluation of myocardial deformation, i.e. strain. Strain² is the deformation produced by the application of a force; myocardial strain represents percent change in myocardial length from relaxed to contractile state. Unlike EF, strain allows studying the different spatial components of contractile function in either longitudinal strain (LS), circumferential strain (CS), or radial strain (RS) directions, both globally and regionally. However, similar to EF, strain

represents a load-dependent estimation of cardiac function and neither is able to depict the true myocardial contractility.

Assessment of LV deformation through quantification of strain has witnessed considerable development, from echocardiographic determined velocity of circumferential fibre shortening,³ cardiac magnetic resonance (cMR) tissue tagging,⁴ tissue Doppler echocardiography^{5–8} to current speckle tracking echocardiography (STE), and feature tracking (FT) approaches.^{9–12} Alterations of strain were found to occur in the setting of maintained EF^{13,14} and were reported to provide additional prognostic value over EF alone in a multitude of clinical scenarios, ranging from asymptomatic adults without a previous history of cardiac pathology (as participants of MESA and Framingham studies)^{15,16} to valvular heart disease (in particular aortic stenosis¹³) and heart failure with preserved and reduced EF.^{17,18} Therefore, deformation imaging techniques have become extremely popular and,

* Corresponding author. Tel: +32 (2) 764 2803; Fax: +32 (2) 764 8980. E-mail: Bernhard.gerber@uclouvain.be

© The Author(s) 2019. Published by Oxford University Press on behalf of the European Society of Cardiology. This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/4.0/>), which permits non-commercial re-use, distribution, and reproduction in any medium, provided the original work is properly cited. For commercial re-use, please contact journals.permissions@oup.com

being applied to numerous research questions, have resulted in an extensive number of published papers, with 'myocardial strain' keyword search hitting nearly 8000 results in PubMed alone. Notwithstanding the enthusiastic scientific interest, myocardial deformation assessment has only partly breached the clinical setting, as several concerns have been raised regarding its robustness in the real-life scenario.

In this review, we attempt to summarize the general principles and technical particularities of current deformation imaging modalities, with particular emphasis on factors explaining differences in measurement values among methods. Further, we aim to provide an overview of current state of validation and intra- and inter-modality comparison.

General principles of deformation imaging techniques

Myocardial deformation can be assessed both from echocardiographic and cMR images, following a similar general workflow, with specific analysis algorithms implemented for each imaging modality. Most deformation imaging techniques share the common principle that specific patterns or features are identified within an image and followed over time in the subsequent images of the sequence by searching the most probable correspondence in successive image frames.¹⁹ Then, local tissue deformation can be estimated by repeating the process for the entire time sequence.

Tissue tracking—general workflow


Typical workflow of tissue tracking is shown in Figure 1. Generally, the *initial step* is to recognize the key cardiac events: end-diastole (ED) and end-systole (ES). The *second step* is the definition of a region of interest encompassing the myocardial wall, by semi-automatic contouring of the endocardial and epicardial borders either in ED or in ES or both. *Segmentation* is a critical step as it defines the set of points that will be tracked, introducing variability depending on the user and the segmentation algorithm.¹⁰ Finally, the region of interest is *tracked* throughout the cardiac cycle, and *strain curves* are computed, possibly post-processed. Either the end-systolic or peak systolic strain can be reported.

Technology of tissue tracking—analysis algorithms

Echocardiographic and cMR deformation imaging softwares employ different algorithms to process the image in order to estimate the local myocardial motion. Some techniques exploit specificities of the imaging modality [e.g. cMR tagging], while others are generic and can be applied to any modality (e.g. block-matching techniques for STE and FT). A more detailed technical discussion is included in the [Supplementary data](#) online.

Specific modalities of tissue tracking and strain imaging

The tissue tracking strategies can be applied to echocardiographic (2D or 3D), cMR (cine or tagged images), sometimes extending these strategies to account for specificities of the imaging modality



(Figure 2). The particularities of these techniques to estimate myocardial deformation are discussed below, while advantages and shortcomings of each method are summarized in the following sections.

cMR tagging

cMR tagging¹ magnetically labels different regions in the myocardium, by creating, prior to image acquisition, locally induced perturbations of the magnetization with selective radiofrequency saturation planes²⁰ resulting in dark lines. When the saturation pulses are applied in two orthogonal planes, the resulting tagging pattern forms a grid of intrinsic tissue markers known as tags. Because the magnetization is a property of the tissue, the tag lines move along with the tissue in which they are created, deforming during contraction. Tag using harmonic phase imaging (HARP) technique will find the best optical flow for matching the multiple 'channels' of the tagged acquisitions, each tag direction corresponding to one channel. Thus, tracking the tag deformation allows direct evaluation of the myocardial deformation or strain. Variations of tagging for strain computation are Strain Encoding magnetic resonance imaging (SENC)²¹ and Displacement Encoding with Stimulated Echos (DENSE).²² In these techniques, encoding is applied through plane and pixel intensities directly relate to the amount of tissue deformation.

cMR tagging has been widely accepted as the reference standard imaging modality for strain quantification after extensive validation *in vitro*²³ and *in vivo*^{24–31} and has allowed the development of the first models of normal and abnormal myocardial motion in humans.^{24,29,32–35} The main advantage of tagging is that deformation is directly measured by physical properties of the tissue. Yet, cMR tagging also has certain limitations (Table 1). Tagged images have low




Figure 2 The principles of tissue tracking techniques illustrated on different imaging modalities. The myocardial speckled pattern (on 2D and 3D echocardiography), or anatomical features (on cine-cMR images), or tagging information (on cMR tagging) are identified within an image and followed over time in the subsequent images of the sequence by searching the most probable pattern correspondence.

temporal resolution reaching at the best 20–30 frames/heart-beat. Furthermore, tag deposition in the beginning of systole starts after detection of R wave and introduces a delay of approximately 30 ms. Thus tag deposition may not be exactly at the beginning of cardiac contraction, potentially leading to underestimation of strain, especially at high heart rates. The spatial resolution of tags, as well as the ratio of tag spacing to slice thickness, are also important factors for reliable strain measurements. For this reason, the accuracy of strain estimates from cMR tagging is lower at the endocardial border and in thin-walled regions of the LV, and cMR tagging estimates essentially mid-wall rather than endocardial strain. Finally, tagging requires dedicated acquisition sequence and time-consuming post-processing using specific software solutions such as HARP. Therefore, cMR tagging has mainly remained a research tool and has not undergone as widespread use as more recent methods to measure strain.

Speckle tracking echocardiography

STE is currently the widest available technique to quantify myocardial deformation,³⁶ mainly because it can be performed on conventional B-Mode images, assuming that image quality is sufficient.

STE analyses LV deformation by tracking cardiac motion from image intensities. Features being tracked can include image contours and image texture, more specifically, the naturally occurring speckled pattern of the myocardium when imaged by ultrasound.³⁷ For tracking the speckle texture, block-matching method is a commonly used technique. It automatically identifies a pattern within a region or block of interest, compares it to all possible matching regions within the search region and finds the position of the best matching block compared with the original one. STE can be applied to 2D, and more recently to 3D echocardiographic images. Optical flow methods have also been applied to echocardiographic images, as well as elastic

Table 1 Spatial and temporal resolution and strength and weaknesses of different imaging modalities

	2DSTE	3DSTE	cMR-FT	cMR Tagging	cMR SENC	cMR DENSE
Spatial	0.2–0.3 mm	0.4–0.5 mm	1–2 mm in plane 6–10 mm through plane	>1 mm in plane 5–7 mm through plane	1.5–2 mm	1.5–2 mm
Temporal	40–60 frames/s	20–50 frames/s	25–35 phases/heart-beat	20–30 phases/heart-beat	20–30 phases/heart-beat	20–30 phases/heart-beat
Strengths	• Ease and availability • High temporal and spatial resolution	• Good reproducibility of planes • Less foreshortening • Better for CS	• Ease (analysis on standard SSFP cine images) • Reproducibility of planes • Several commercial softwares • Good for LS, CS, and RS	• True tissue markers • Extensive validation • Higher reproducibility of plane acquisitions • 2D and 3D (three strain directions)	• High spatial resolution • Short acquisition time (1 heart-beat) • Fast post-processing • Allows real-time strain for stress cMR	• High spatial resolution • Fast post-processing • Three strain directions form 2D acquisitions
Weaknesses	• Foreshortening • Reproducibility of acquisition planes, particularly for CS and rotation/twist • Through-plane motion • Less performant for CS, RS, and regional strains	• Lower spatial and temporal resolution than 2D STE • Less available than 2D STE • Multibeam acquisition with limited temporal resolution when arrhythmia	• No physical speckles or intra-tissue markers (based on contours only) • Less performant for regional strain • Low spatial and temporal resolution • Less validated • 2D strains only • No rotation/twist	• Requires special sequences and analysis software • Few commercial softwares • Time-consuming acquisition and analysis • Through-plane motion of tags • Tag fading in diastole • Low spatial and temporal resolution • Tag deposition delay may lead to underestimation of strain	• Mainly a research technique • Requires special sequences and analysis software • Not 3D • Low temporal resolution • Measures only through plane strain (CS from long axis, LS from short axis) • No radial strain	• Mainly a research technique • Requires special sequences and analysis software • Low temporal resolution • No true 3D

registration, all of them being able to capture motion and to a certain extent deformation as demonstrated on synthetic images.³⁸ STE has high spatial and temporal resolution, but depending on the algorithm, typically evaluates speckle motion mainly at endocardial border of the LV, and relatively less in the myocardium.

cMR-FT

cMR-FT is a relatively new 2D imaging technique that can be applied to standard cMR cine SSFP sequences, gaining popularity by allowing measurement of myocardial deformation without the need for dedicated acquisition and complex post-processing.^{12,19}

cMR-FT is mainly based on a block-matching approach. It first identifies anatomic features in the cMR image along the myocardial boundaries, defines region of interests around these locations and track them along the cardiac cycle by looking for the most similar region in

the next image. Advantages of FT is that strain can be computed on conventional SSFP cine images using several commercial softwares. In contrast to STE and cMR tagging, FT does not seem to distinguish intramyocardial features, as the grey level distribution in cine SSFP images is relatively homogenous. Furthermore, similar to tagging, cine-cMR images have substantially lower spatial and temporal resolution than STE (Table 1).

Sources of variations and intra- and inter-modality inconsistencies

The differences in imaging modalities and competitive methodologies result in intra- and inter-modality inconsistencies in deformation estimation as illustrated by Figure 3A and B. These are explained by several factors, as listed below and summed up in Table 2.




Figure 3 Example of differences in global and regional strain estimates (A longitudinal and B circumferential) by different modalities and softwares in a patient with hypertrophic cardiomyopathy. Regional strain values are represented in 17 and 18 segment colour-coded bullseyes plots. Excluded segments due to poor image quality or tracking are not colour-coded.

Table 2 Summary of sources of variations and intra- and inter-modality inconsistencies

Imaging modality related factors	Quality of the acquisition process
	Spatial and temporal resolution
	Segmentation misalignment between imaging modalities
Software-related factors	Spatial and temporal smoothing
	Size of the search region
	Favouring tracking in a certain myocardial layer
	Computation of Lagrangian or Eulerian strain
	Calculation of global strain values
Operator-related factors	Definition of end-diastole and end-systole
	Definition of regions of interest
	Experience and training

Imaging modality related factors

A first set of factors potentially influencing deformation quantification is the *quality of the acquisition process*, varying between operators and modalities. Ensuring reproducible and accurate breathing control is therefore a key requirement for all modalities.

A second set of modality-dependent factors relate to the *spatial and temporal resolution* of the images. Both resolutions are crucial to ensure the complete characterization of myocardial deformation over successive time frames. If the temporal and spatial resolution is too low (*Figure 4*), the local patterns may become less comparable, an effect known as image de-correlation and displacements may become harder to detect.¹⁹ Precisely, the temporal resolution of cMR and 3DSTE is lower than that of 2DSTE, and inferior to Tissue Doppler Imaging for example, meaning that cMR and 3DSTE are more prone to miss the short-lived events during the isovolumic period. Another parameter that may influence strain values is the reference method. For instance in tagged images, tag deposition is delayed relative to electrocardiogram signal detection, which leads to underestimation of strain (*Figure 5*). On the other hand, STE has a higher spatial resolution than cMR, which is however blunted by a low signal-to-noise ratio. The particularity of the spatial resolution of the ultrasound images is a lower lateral than axial resolution and lower in-depth resolution.³⁹ This means that most reliable results are obtained closer to the centre line of the image, at smaller depths.

Finally, a last source of potential discrepancies in 2D regional deformation values between modalities is the difficulty to match myocardial segments. This *segmentation misalignment between imaging modalities* is explained by the fact that 2D imaging planes are not necessarily the same when acquired by echocardiography and cMR due to different scanning angles, therefore, complete correspondence between segments is not achievable. For any 2D technique, the pattern within a region of interest is detected and tracked along the image plane. However, as LV deformation is a 3D phenomenon, involving a combination of apex-to-base shortening and simultaneous twisting, the myocardial patterns have a complex 3D motion. Therefore, the pattern within a region of interest defined in a 2D image plane, might




Figure 4 Influence of temporal resolution on strain measurement (data extrapolated from a high temporal resolution STE image undersampled at lower frame rates).

move out of the scanning plane during the cardiac cycle. Moreover, unlike cMR, the imaging planes of 2D echo might not depict the true apex in long- and short-axis views, effect known as foreshortening. The through-plane motion and foreshortening represent a limitation of 2D analysis, which are overcome by 3D techniques.

Software-related factors

Several factors related to the specificities of implementing image tracking algorithms can also heavily influence deformation values.

ST algorithms apply *spatial and temporal smoothing* to regularize the results in order to reduce noise, which can affect the measurement robustness by missing significant localized abnormalities in the case of spatial smoothing or by masking rapid events in case of temporal smoothing.^{19,40} Also, as strain is computed from the spatial derivatives of the displacement, different regularization strategies (thus affecting motion smoothness) can dramatically affect the range of deformation values computed from the displacement field, at least when considering single material points or small regions. Therefore, parameters based on local estimates are more prone to variability than those based on an integrative combination, i.e. global strains are more stable and reliable than segmental strains.

For block-matching algorithms, the *size of the search region* must be carefully tuned.¹⁹ In general, solving for displacements between short distance regions is challenging and may explain why usually RS (computed on the small distance between endo- and epicardium) is less reliable than LS and CS⁴¹ that are computed over larger regions.

Favouring the tracking in a certain myocardial layer, i.e. endocardial rather than transmural could alter the strain values, as given the fibre orientation, the deformation of the endocardial layer is more important than in the mid and epicardial layers. The level of endocardial strain detection by the software is probably the most important factor (*Figure 6*) inducing intra- and inter- modality variability in strain measurement. Importantly, the level of layer detection may also vary among imaging methods. In particular, as mentioned before, tags are

mainly detected in mid-wall, whereas STE mainly follows endocardial markers.

Other software-related factors that introduce variability are:

- *Computation of Lagrangian or Eulerian strain*, as the two formulas (represented in Figure 4) will result in slightly different values, with Eulerian strain having higher absolute values (Figure 7).
- *Calculation of global strain values*, either by using the entire myocardial length or by averaging values computed at segmental level, will give different results.⁴²




Figure 5 Influence of tag deposition delay of strain computation in cMR tagging [data extrapolated from a high resolution STE image acquired at high heart rate (120 bpm)].

- *Definition of ED and ES*, which has been shown to have a major influence on the accuracy of strain measurements, up to the point that changing ED or ES by only four frames can significantly impact strain values to as much as 20–40% relative changes in ES GLS.⁴³

Operator-related factors

Most strain analysis softwares require *manual drawing of the myocardial contours*. As these contours define the regions/points being tracked, different operators contouring differently will obtain different deformation values. For all techniques, operator *experience and training* is an important factor in accuracy of measurements. Indeed most validation studies were performed in highly experienced centres and core-labs and may not translate to overall clinical practice. With the advent of machine learning and fully automated analysis this factor may become less important in the future.

Practical aspects

Tissue tracking software platforms may use different algorithms for measuring deformation and presentation of results, therefore, two aspects become critically important: validation of each specific analysis software and consensus reporting among software package vendors. Validating non-invasive tools used for clinical practice is, however, a challenging task, and has not been done for routine parameters as EF for example. Additionally, EF is subjected to inter-modality (echocardiography vs. cMR) and inter-vendor (different 3DE analysis softwares) variability on top of the suboptimal inter and intra-observer reproducibility.





Figure 6 Influence of endocardial layer position on strain measurements. (Example of different layer positions on endocardial strain in a cMR-FT image).




Figure 7 Difference in Eulerian and Lagrangian strain computation.

Table 3 *In vitro* validation of 2DSTE and 3DSTE

Study	Model	Method	Reference	Software	Strain	Conclusion			
						r/ICC	Bias ± 2 SD (%)	95% CI	
2DSTE									
Korinek et al. ⁴⁴	Phantom	Different motion rates (n = 23)	Sono	GE EchoPAC	Long	r = 0.99	0.7 ± 2.2	-3.6 to 5	Promising
Amzulescu et al. ⁴⁵	Phantom	Different motion rates and stroke volumes (n = 35)	Sono	Qlab 10.3 Philips	Long	ICC = 0.893 ± 2.8		-8.2 to 2.5	Good for Long
3DSTE									
Heyde et al. ⁴⁶	Phantom	Different motion rates (n = 7)	Sono	In-house software	Long	r = 0.92			Adequate
					Circ	r = 0.84			
					Rad	r = 0.96			
Hjertaas et al. ⁴⁷	Phantom	Different motion rates and stroke volumes (n = 15)	Sono	GE EchoPAC	Long	r = 0.99	0.8 ± 1.5	-2.1 to 3.7	Accurate for Long and Circ, not for Rad
				BT11	Circ	r = 0.98	-0.7 ± 1.7	-4 to 2.6	
					Rad	r = 0.89	16.1 ± 22.2	-27.4 to 59.6	

Validation

Before implementation in the clinical setting, any new strain imaging method requires a complex process of validation: on synthetic data sets, *in vitro* and *in vivo* experiments, and validation in humans. In opposition to the enthusiastic number of published papers assessing the clinical usefulness of strain, there is a notable lack of validation studies. Indeed, while numerous studies have performed inter-technique comparisons, which can only demonstrate their relative performance, true validation studies for STE, and cMR-FT studies have only been performed for few commercial softwares of the numerous available alternatives. Additionally, most of the clinical validation

studies have included a very small number of subjects in the healthy control group and even smaller number in the diseased group, which can only hardly represent the diverse pathological phenotypes encountered in clinical practice.

Validation on synthetic data sets

Synthetic data sets are computer generated images and constitute the first step when testing a new software. In such a controlled environment, the deformation values to be measured are known (i.e. ground truth), while parameters as motion rate, wall thickness and EF can be synthetically altered to simulate different cardiac conditions.

Table 4 *In vivo* validation of 2DSTE and 3DSTE

Study	Model	Method	Reference	Software	Strain	Conclusion			
						r/ICC	Bias ± 2 SD (%)	95% CI	
2DSTE									
Korinek et al. ⁴⁴	16 pigs	Baseline, LAD ligation	Sono	GE EchoPAC PC_2D strain	Long	r = 0.94	-1.1 ± 7.5	-15.8 to 3.9	Promising
Toyoda et al. ⁵¹	6 dogs	Dobutamine	Sono	US customized software	Rad	r = 0.92			Promising
Langeland et al. ⁵²	5 sheep	Baseline, CX ligation, esmolol, dobutamine	Sono	In-house software (SPEQLE 2D)	Long	ICC = 0.80	0.4 ± 2.7	-5 to 5.8	Promising
					Rad	ICC = 0.72	2 ± 4.6	-7.1 to 11	
Amundsen et al. ⁵³	9 dogs	Baseline, saline loading, LAD occlusion	Sono	MathLab-based custom made programme	Long	r = 0.9		-4.4 to 5	Accurate
					Rad	r = 0.79		-5.6 to 5.1	
Reant et al. ⁵⁴	10 pigs	Baseline, LAD occlusion, dobutamine	Sono	GE EchoPAC	Long	ICC = 0.93–0.96			Real potential
					Circ	ICC = 0.50–0.73			
					Rad	ICC = 0.98			
Pirat et al. ⁵⁵	7 dogs	Baseline, LAD occlusion, esmolol, dobutamine	Sono	Siemens VVI	Long	r = 0.83–0.90			Accurate
					Circ	r = 0.88–0.94			
Heyde et al. ⁵⁶	5 sheep	Baseline, CX ligation, esmolol, dobutamine	Sono	GE EchoPAC v110.0.0,	Long	r = 0.69	-2.1	-9.5 to 5.3	Circ and Radial
					Circ	r = 0.72	-7.7	-19.5 to 4.1	overestimate
					Rad	r = 0.64	18.5	-4.6 to 41.7	
3DSTE									
Seo et al. ⁵⁷	10 sheep	Baseline, LAD ligation, dobutamine, propranolol	Sono	Toshiba 3D wall motion tracking	Long	r = 0.89			Reliable
					Circ	r = 0.90			
					Rad	r = 0.84			
Heyde et al. ⁵⁸	14 sheep	Baseline, CX ligation, dobutamine, esmolol	Sono	In-house STE software	Long	r = 0.64			Acceptable
					Circ	r = 0.62			accuracy
					Rad	r = 0.69			
Bouchez et al. ⁵⁹	13 sheep	Baseline, dobutamine, CX occlusion	Sono	Siemens eSie volume mechanics	Long	r = 0.78	-5 ± 6		Good for Long
					Circ	r = 0.71	-5 ± 7		and Circ, less
					Rad	r = 0.30	15 ± 19		accurate for Rad

Currently, open-access libraries of 2D and 3D simulated ultrasound datasets, as well as simulated cine cMR, are being built to facilitate performance analysis of different software packages in order to promote quality assurance.^{48–50} The tested strain imaging methods have shown promising results, and efforts have been made to reach the level of realism of the real ultrasound and cMR images.

In vitro validation

The next step used for validation of strain techniques is by using physical cardiac phantoms in which motion is mechanically controlled. In this case, the motion of the phantom is compared to the ground truth obtained by sonomicrometry recordings. Sonomicrometry is a technique of measuring distances between piezoelectric crystals based on the speed of acoustic signals through the medium they are embedded in. Both 2DST⁴⁴ and 3DST^{46,47} methods have been validated *in vitro*

and have shown good accuracy (see Table 3), while for FT there is currently no validation on phantoms.

However, the models used to mimic motion are generally simple and do not represent the true complex cardiac deformation, while cardiac anatomic structures as trabeculations/valves are not represented.

In vivo validation

In order to approximate the real-life conditions, an *in vivo* design to validate strain measurements is required. Different open-chest animal models have been used and myocardial deformation values have been compared to sonomicrometry. Studies investigating 2DSTE and 3DSTE (Table 4) have reported overall good agreement of strain by STE with sonomicrometry measurements. Generally, while LS 2DSTE seems to perform well across studies, recent reports describe suboptimal correlation and larger bias for CS and RS by




Figure 8 Reported normal (mean and 95% confidence interval) global strain values in healthy subjects for different imaging modalities. Data from refs. ^{45,65,68,82–91} Normal GLS, GCS, and GRS values were compared using random effects models weighted by inverse variance and heterogeneity between methods was compared using the Cochran Q test and the inconsistency factor. For all strain measurements, I^2 and Q indicated significant heterogeneity among studies and methods.

remains largely neglected, is the lack of proper validation of most methods vs. absolute and objective reference standard.

To allow accurate deformation estimates and avoid unnecessary variability between products and methods, it should be mandatory that each strain quantification method undergoes rigorous validation using a multi-step process before wide-use for research purposes, and even more, for clinical implementation. Despite the difficulties, such an approach of widespread validation and cross-modality and vendor standardization needs to be applied to allow further development of this technology and successful clinical utilization of these methods.

Supplementary data

Supplementary data are available at *European Heart Journal - Cardiovascular Imaging* online.

Funding

Grant support by the Fondation Nationale de la Recherche Scientifique of the Belgian Government (FRSM PDR 19488731).

Conflict of interest: H.L. and M.D.C. are employed by Philips Medical Systems. The Cliniques Universitaires St. Luc have a master research agreement with Philips Medical Systems. All remaining authors have declared no conflicts of interest.

References

- Konstam MA, Abboud FM. Ejection fraction: misunderstood and overrated (changing the paradigm in categorizing heart failure). *Circulation* 2017;**135**: 717–9.
- Mirsky I, Parmley WW. Assessment of passive elastic stiffness for isolated heart muscle and the intact heart. *Circ Res* 1973;**33**:233–43.
- Domanski MJ, Follmann D, Mirsky II. A new approach to assessing regional and global myocardial contractility. *Echocardiography* 1997;**14**:1–8.
- Zerhouni EA, Parish DM, Rogers WJ, Yang A, Shapiro EP. Human heart: tagging with MR imaging—a method for noninvasive assessment of myocardial motion. *Radiology* 1988;**169**:59–63.
- Sutherland GR, Stewart MJ, Groundstroem KW, Moran CM, Fleming A, Guell-Peris FJ. Color Doppler myocardial imaging: a new technique for the assessment of myocardial function. *J Am Soc Echocardiogr* 1994;**7**:441–58.
- Heimdal A, Støylen A, Torp H, Skjærpe T. Real-time strain rate imaging of the left ventricle by ultrasound. *J Am Soc Echocardiogr* 1998;**11**:1013–9.
- Edwardsen T, Gerber BL, Garot J, Bluemke DA, Lima JA, Smiseth OA. Quantitative assessment of intrinsic regional myocardial deformation by Doppler strain rate echocardiography in humans: validation against three-dimensional tagged magnetic resonance imaging. *Circulation* 2002;**106**:50–6.
- Urheim S, Edvardsen T, Torp H, Angelsen B, Smiseth OA. Myocardial strain by Doppler echocardiography. Validation of a new method to quantify regional myocardial function. *Circulation* 2000;**102**:1158–64.
- Collier P, Phelan D, Klein A. A test in context: myocardial strain measured by speckle-tracking echocardiography. *J Am Coll Cardiol* 2017;**69**:1043–56.
- Jasaityte R, Heyde B, D'Hooge J. Current state of three-dimensional myocardial strain estimation using echocardiography. *J Am Soc Echocardiogr* 2013;**26**:15–28.
- Muraru D, Niero A, Rodriguez-Zanella H, Cherata D, Badano L. Three-dimensional speckle-tracking echocardiography: benefits and limitations of integrating myocardial mechanics with three-dimensional imaging. *Cardiovasc Diagn Ther* 2018;**8**:101–17.
- Claus P, Omar AMS, Pedrizzetti G, Sengupta PP, Nagel E. Tissue tracking technology for assessing cardiac mechanics: principles, normal values, and clinical applications. *JACC Cardiovasc Imaging* 2015;**8**:1444–60.
- Delgado V, Tops LF, van Bommel RJ, van der Kley F, Marsan NA, Klautz RJ et al. Strain analysis in patients with severe aortic stenosis and preserved left ventricular ejection fraction undergoing surgical valve replacement. *Eur Heart J* 2009;**30**: 3037–47.
- Kraigher-Krainer E, Shah AM, Gupta DK, Santos A, Claggett B, Pieske B et al. Impaired systolic function by strain imaging in heart failure with preserved ejection fraction. *J Am Coll Cardiol* 2014;**63**:447–56.
- Choi E-Y, Rosen BD, Fernandes VRS, Yan RT, Yoneyama K, Donekal S et al. Prognostic value of myocardial circumferential strain for incident heart failure and cardiovascular events in asymptomatic individuals: the Multi-Ethnic Study of Atherosclerosis. *Eur Heart J* 2013;**34**:2354–61.
- Cheng S, McCabe EL, Larson MG, Merz AA, Osypuk E, Lehman BT et al. Distinct aspects of left ventricular mechanical function are differentially associated with cardiovascular outcomes and all-cause mortality in the community. *J Am Heart Assoc* 2015;**4**:e002071.
- Park JJ, Park JB, Park JH, Cho GY. Global longitudinal strain to predict mortality in patients with acute heart failure. *J Am Coll Cardiol* 2018;**71**:1947–57.
- Shah AM, Claggett B, Sweitzer NK, Shah SJ, Anand IS, Liu L et al. Prognostic importance of impaired systolic function in heart failure with preserved ejection fraction and the impact of spironolactone. *Circulation* 2015;**132**:402–14.
- Pedrizzetti G, Claus P, Kilner PJ, Nagel E. Principles of cardiovascular magnetic resonance feature tracking and echocardiographic speckle tracking for informed clinical use. *J Cardiovasc Magn Reson* 2016;**18**:51.
- Axel L, Dougherty L. MR imaging of motion with spatial modulation of magnetization. *Radiology* 1989;**171**:841–5.
- Osman NF, Sampath S, Atalar E, Prince JL. Imaging longitudinal cardiac strain on short-axis images using strain-encoded MRI. *Magn Reson Med* 2001;**46**:324–34.
- Kim D, Gilson WD, Kramer CM, Epstein FH. Myocardial tissue tracking with two-dimensional cine displacement-encoded MR imaging: development and initial evaluation. *Radiology* 2004;**230**:862–71.
- Young AA, Axel L, Dougherty L, Bogen DK, Parenteau CS. Validation of tagging with MR imaging to estimate material deformation. *Radiology* 1993;**188**:101–8.
- Garot J, Bluemke DA, Osman NF, Rochitte CE, McVeigh ER, Zerhouni EA et al. Fast determination of regional myocardial strain fields from tagged cardiac images using harmonic phase MRI. *Circulation* 2000;**101**:981–8.
- Yeon SB, Reichel N, Tallant BA, Lima JA, Calhoun LP, Clark NR et al. Validation of *in vivo* myocardial strain measurement by magnetic resonance tagging with sonomicrometry. *J Am Coll Cardiol* 2001;**38**:555–61.
- Lima JA, Jeremy R, Guier W, Bouton S, Zerhouni EA, McVeigh E et al. Accurate systolic wall thickening by nuclear magnetic resonance imaging with tissue tagging: correlation with sonomicrometers in normal and ischemic myocardium. *J Am Coll Cardiol* 1993;**21**:1741–51.
- Thomas D, Ferrari VA, Janik M, Kim DH, Pickup S, Glickson JD et al. Quantitative assessment of regional myocardial function in a rat model of myocardial infarction using tagged MRI. *MAGMA* 2004;**17**:179–87.
- Liu W, Chen J, Ji S, Allen JS, Bayly PV, Wickline SA et al. Harmonic phase MR tagging for direct quantification of Lagrangian strain in rat hearts after myocardial infarction. *Magn Reson Med* 2004;**52**:1282–90.
- McVeigh ER, Zerhouni EA. Noninvasive measurement of transmural gradients in myocardial strain with MR imaging. *Radiology* 1991;**180**:677–83.
- Zhou R, Pickup S, Glickson JD, Scott CH, Ferrari VA. Assessment of global and regional myocardial function in the mouse using cine and tagged MRI. *Magn Reson Med* 2003;**49**:760–4.
- Azhari H, Weiss JL, Rogers WJ, Siu CO, Shapiro EP. A noninvasive comparative study of myocardial strains in ischemic canine hearts using tagged MRI in 3-D. *Am J Physiol* 1995;**268**:H1918–26.
- Clark NR, Reichel N, Bergey P, Hoffman EA, Brownson D, Palmon L et al. Circumferential myocardial shortening in the normal human left ventricle. Assessment by magnetic resonance imaging using spatial modulation of magnetization. *Circulation* 1991;**84**:67–74.
- Moore CC, McVeigh ER, Zerhouni EA. Quantitative tagged magnetic resonance imaging of the normal human left ventricle. *Top Magn Reson Imaging* 2000;**11**: 359–71.
- McVeigh ER. MRI of myocardial function: motion tracking techniques. *Magn Reson Imaging* 1996;**14**:137–50.
- Young AA, Axel L. Three-dimensional motion and deformation of the heart wall: estimation with spatial modulation of magnetization—a model-based approach. *Radiology* 1992;**185**:241–7.
- Geyer H, Caracciolo G, Abe H, Wilansky S, Carerj S, Gentile F et al. Assessment of myocardial mechanics using speckle tracking echocardiography: fundamentals and clinical applications. *J Am Soc Echocardiogr* 2010;**23**:351–69; quiz 453–5.
- Perk G, Tunick PA, Kronzon I. Non-Doppler two-dimensional strain imaging by echocardiography—from technical considerations to clinical applications. *J Am Soc Echocardiogr* 2007;**20**:234–43.
- De Craene M, Marchesseau S, Heyde B, Gao H, Alessandrini M, Bernard O et al. 3D strain assessment in ultrasound (Straus): a synthetic comparison of five tracking methodologies. *IEEE Trans Med Imaging* 2013;**32**:1632–46.
- Smiseth OA, Torp H, Opdahl A, Haugaa KH, Urheim S. Myocardial strain imaging: how useful is it in clinical decision making? *Eur Heart J* 2016;**37**:1196–207.
- Bijnens BH, Cikes M, Claus P, Sutherland GR. Velocity and deformation imaging for the assessment of myocardial dysfunction. *Eur J Echocardiogr* 2009;**10**: 216–26.
- Langeland S, Wouters PF, Claus P, Leather HA, Bijnens B, Sutherland GR et al. Experimental assessment of a new research tool for the estimation of two-dimensional myocardial strain. *Ultrasound Med Biol* 2006;**32**:1509–13.

83. Muraru D, Cucchini U, Mihăilă S, Miglioranza MH, Aruta P, Cavalli G et al. Left ventricular myocardial strain by three-dimensional speckle-tracking echocardiography in healthy subjects: reference values and analysis of their physiologic and technical determinants. *J Am Soc Echocardiogr* 2014;**27**:858–71.e1.
84. Cheng S, Larson MG, McCabe EL, Osypiuk E, Lehman BT, Stanchev P et al. Age- and sex-based reference limits and clinical correlates of myocardial strain and synchrony: the Framingham Heart Study. *Circ Cardiovasc Imaging* 2013;**6**: 692–9.
85. Yingchoncharoen T, Agarwal S, Popovic ZB, Marwick TH. Normal ranges of left ventricular strain: a meta-analysis. *J Am Soc Echocardiogr* 2013;**26**:185–91.
86. Kocabay G, Muraru D, Peluso D, Cucchini U, Mihaila S, Padayattil-Jose S et al. Normal left ventricular mechanics by two-dimensional speckle-tracking echocardiography: Reference values in healthy adults. *Rev Esp Cardiol (Engl Ed)* 2014;**67**: 651–8.
87. Bernard A, Addetia K, Dulgheru R, Caballero L, Sugimoto T, Akhaladze N et al. 3D echocardiographic reference ranges for normal left ventricular volumes and strain: results from the EACVI NORRE study. *Eur Heart J Cardiovasc Imaging* 2017;**18**:475–83.
88. Kleijn SA, Pandian NG, Thomas JD, Perez de Isla L, Kamp O, Zuber M et al. Normal reference values of left ventricular strain using three-dimensional speckle tracking echocardiography: results from a multicentre study. *Eur Heart J Cardiovasc Imaging* 2015;**16**:410–6.
89. Vo HQ, Marwick TH, Negishi K. MRI-derived myocardial strain measures in normal subjects. *JACC Cardiovasc Imaging* 2018;**11**:196–205.
90. Venkatesh BA, Donekal S, Yoneyama K, Wu C, Fernandes VR, Rosen BD et al. Regional myocardial functional patterns: quantitative tagged magnetic resonance imaging in an adult population free of cardiovascular risk factors: the multi-ethnic study of atherosclerosis (MESA). *J Magn Reson Imaging* 2015;**42**:153–9.
91. Giusca S, Korosoglou G, Zieschang V, Stoiber L, Schnackenburg B, Stehning C et al. Reproducibility study on myocardial strain assessment using fast-SENC cardiac magnetic resonance imaging. *Sci Rep* 2018;**8**:14100.
92. Andre F, Steen H, Matheis P, Westkott M, Breuninger K, Sander Y et al. Age- and gender-related normal left ventricular deformation assessed by cardiovascular magnetic resonance feature tracking. *J Cardiovasc Magn Reson* 2015;**17**:25.
93. Knackstedt C, Bekkers SC, Schummers G, Schreckenber M, Muraru D, Badano LP et al. Fully automated versus standard tracking of left ventricular ejection fraction and longitudinal strain: the FAST-EFs Multicenter Study. *J Am Coll Cardiol* 2015;**66**:1456–66.
94. Medvedofsky D, Kebed K, Laffin L, Stone J, Addetia K, Lang RM et al. Reproducibility and experience dependence of echocardiographic indices of left ventricular function: side-by-side comparison of global longitudinal strain and ejection fraction. *Echocardiography* 2017;**34**:365–70.
95. Barbier P, Mirea O, Cefalu C, Maltagliati A, Savioli G, Guglielmo M. Reliability and feasibility of longitudinal AFI global and segmental strain compared with 2D left ventricular volumes and ejection fraction: intra- and inter-operator, test-retest, and inter-cycle reproducibility. *Eur Heart J Cardiovasc Imaging* 2015;**16**:642–52.
96. Reichek N. Myocardial strain: still a long way to go. *Circ Cardiovasc Imaging* 2017; **10**:e007145.